

**Minutes of the Pre-bid Meeting for Procurement of Rio Grande ADCP, held on  
14.02.2020, 11.30 AM, in the Conference Hall, 1<sup>st</sup> Floor, B2 Block, NCSCM, Chennai**  
[Package No. NPMU/G-112]

**Members Present:**

1. Dr. D. Chandramohan, External Expert
2. Dr. Dipnarayan Ganguly, Scientist C
3. Dr. Gurmeet Singh, Scientist C
4. Dr. Robin R. S., Scientist C
5. Mr. Alok Ranjan Samal, Manager - Finance/ Accounts
6. Mr. Pradeep Tripathy, Procurement Consultant

**Representatives of Participating Firms:**

1. M/s. Elektroniklab India Pvt. Ltd. - Mr. Sivakumar D.
2. M/s. Elektroniklab India Pvt. Ltd. - Mr. Ramesh
3. M/s. Elektroniklab India Pvt. Ltd. - Mr. Suresh

As scheduled, the pre-bid meeting for procurement of Rio Grande ADCP was held on 14.02.2020 under the chairmanship of Dr. Chandramohan, External Expert, NCSCM, Chennai. Only one prospective bidder attended the pre-bid meeting through three nos. of representatives. The Chairman welcomed all the participants to the meeting.

The participants were invited to present their queries on the bid document for Supply of Rio Grande ADCP. The queries/ suggestions made by the prospective bidders and the clarification/ response thereon along with the modification in the relevant provisions of the bid document are appended here with as **Annex-A**.

The prospective bidder was invited to participate in the bidding process after careful consideration to provisions of the bid document published and possible amendment/s subsequent to the present pre-bid meeting.

**Annex- A**

**Queries/ Suggestions and response thereon for Procurement of Rio Grande ADCP**

Sl. No.	Particulars	Query / Suggestion	Clarification / Response
1.	Name of the Equipment : RIO GRANDE ADCP	The title of the ADCP states RIO GRANDE ADCP, this title is not clear to us and hence we request you to make the title generic.	The name of the Equipment be read as “ADCP”.
2.	Type of measurement	Please clarify whether the ADCP instrument required is only for current measurement or whether bathymetry function is also required?	The ADCP instrument is required both for current measurement as well as bathymetry function.  <b>Please refer to Amendment.</b>

**Amendment to Technical Specifications**

Particulars as per the published Bid Document (Page No.62 of the Bid Document)		Modification now made	
Upgradable	The complete system should be easily upgradable for Bathymetry application in the future without any hardware change.	River Discharge Applications as well as Bathymetry Survey and Velocity Mapping Survey	Necessary softwares & licenses for both River Discharge Applications as well as Bathymetry Survey & Velocity Mapping Survey should be supplied along with the instrument.

**Note: The amendment mentioned above is to be treated as amendment to the terms and conditions of the bid document. All other terms & conditions of the published bid document shall remain the same.**

**The bidders have to submit a copy of this Minutes along with the amendment and bid documents. Each page of the bid document as well as this Minutes along with the amendment are to be signed by the authorized Signatories of the bidders.**

Sd/-  
**DIRECTOR**